

DOD Medical Materiel Quality Control Messages (MMQC)/Medical Materiel Information Messages (MMI) & Images

Medical Material Quality Control Message

MMQC-14-1652

Message Number: MMQC-14-1652
Released On: 22-JUL-14
Message Class: Informational
Material Type: MedSurg
MMQC\MMI Reference:
FDA Recall Reference:
ECRI Reference:
DLA Reference: DLA-14-296A
MFG Release Date:

Subject:

Remove & destroy all First Aid Kit, Eye Dressing from all Service First Aid Kits

Reason / Information:

The Joint Trauma System (JTS) Clinical Practice Guidelines, Tactical Combat Casualty Care Guidelines, and the Vision Center of Excellence's recommendations advocate the use of a rigid eye shield and rapid evacuation to an eye care provider when treating traumatic eye injuries. This is the only authorized clinical practice guideline for treating traumatic eye injuries (See Image on Message under Additional Documentation/Attachment). Reference NSN directs First Responders to instill ointment into injured eye and apply pressure patch. This is a direct contraindication to the approved clinical practice guidelines and may result in further injury and/or permanent loss of eyesight.

Disposition / Instructions:

Immediately discontinue use, remove, and destroy Reference NSN. Replace reference NSN with NSN 6515-01-598-1877.

DLA T-S Medical is suspending all stocks from issue.

Implementation plan, including specific item disposition instructions, is a service responsibility.

For general questions call 215-737-2891, DSN 444-2891 or by email at James.Johanson@dla.mil.

Note: Protocol directing First Responders to instill ointment into injured eye and apply pressure patch shall be discontinued.

Item Information:

NSN (FSC-NIIN): 6545 - 008536309
Nomenclature: FIRST AID KIT EYE DRESSING
UI: EA
Manufacturer: ALL
Distributor:

LOT / SN:

ALL

ARMY Unit Assemblages

UA	NIIN	NOMENCLATURE
0246	011419470	SUR INSTR&SUP SE INDI
0259	011918972	MES TRAU FLD (1)
0264	012281887	MES TRAUMA FIELD (2)
1352	014184495	RSL DMSO HVY >=36 MO
1360	014370739	MRSL AREA SUPPORT BN
2259	014964835	MES TRAUMA FLD 1
7055	009196650	FIRST AID KIT GE
7061	009221200	FIRST AID KIT GENER
7142	001161410	FIRST AID KIT GEN PUR
7175	010948412	FIRST AID KIT INDIV

Service/Additional Instructions:**AIR FORCE:**

This recall on 6545008536309, First Aid Kit Eye Dressing affects the following assemblages: ACC - 913J V2 AMC - 887C V2

POC Contact Information:**US AIR FORCE:**

Email: afmoa.sgalc@us.af.mil
Phone (COMM): 301-619-4170
Phone (DSN): 343-4170
Fax: 301-619-2557

US ARMY:

Email: usarmy.detrick.medcom-usamma.mbx.doc@mail.mil
Phone (COMM): 301-619-4300
Phone (DSN): 343-4300
Fax: 301-619-4468

US NAVY:

Email: MMQC2@MED.NAVY.MIL

Phone (COMM): 301-619-3085

Phone (DSN): 343-3085

Fax: 301-619-2473

USAMMC-E:

Email: usarmy.rheinland-pfalz.medcom-usammce.list.dtd-quality-assurance@mail.mil

Phone (COMM): 011-49-6331-86-7118/7181

Phone (DSN): 495-7118

Fax: 6218

Additional Message Recipients:

Please ensure dissemination of this information to the following:

Command Channels

Medical Log Officers

Medical Staff

Pharmacy Officers

Supply Officers

Supported Activities/Centers

Message Dissemination Authorization

AF Activities Will Take Action As Prescribed In Afi 41-209, Medical Logistics Support, Chapters 3 And 9. For Majcoms & Ngb--This Msg Has Been Transmitted To All Designated Subordinate Medical Activities.

See Army Regulation (Ar) 40-61, 28 January 2005, Chapter 4, And The Department Of The Army Supply Bulletin (Sb 8-75-11) For Applicable Policies And Procedures.

DODD 5105.22 and DODD 6025.13

Additional Documentation / Attachments

[MMQC-14-1652_1.pdf](#)

*** END OF MESSAGE ***